



199 Fire Tower Drive
Tonawanda, NY USA

Phone: (716) 629-3800
FAX: (716) 693-9162



Certificate of Compliance

The following models, with the TF option:

521, 522

Were evaluated by:

FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102
Norwood, MA 02062, USA

They were found to comply with the following United States Regulations:

FM Class 3600: 2018
FM Class 3610: 2015
FM Class 3810: 2005
ANSI/ISA 61010-1: 2004
ANSI/ISA 60079-0: 2013
ANSI/ISA 60079-11: 2014

This shows that these products are Intrinsically Safe and can be used in the following Hazardous Locations:

Class I, Division 1, Groups A, B, C, and D

Class I, Zone 0 AEx ia IIC

Temperature Code T4 Ta = -40 to 85 °C

Temperature Code T5 Ta = -40 to 40 °C

When installed per the following drawings: [CD0676](#)

This is specified in detail on:

FM Certificate of Conformity: [FM16US0170](#)

The products are marked with the FM Mark and the above ratings

I declare that the above statements are factual.

Joel Neri
Viatran Quality/Approvals Manager

4/9/2021

Date